

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time
Jan Thomanek	04.01.2021	14:17:01

Basic information

Operation parameter	
Script	Atemwiderstand
Aerosol	Pressurized air

Mask information

Mask ID	13868_20210104_141701
Mask los	M04-400010-29.12.2020-002
Mask type	TEST 1 / PALAS / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.54 l/min	63.18 Pa	70.00 Pa	✔ pass
Inhalation 2	93.98 l/min	158.68 Pa	240.00 Pa	✔ pass
Exhalation	159.10 l/min	263.98 Pa	300.00 Pa	✔ pass



Calibration valid until 05.11.2021

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time
Jan Thomanek	04.01.2021	14:21:41

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	93.99 l/min
Generator air	2.97 l/min

Test conditions	
Temperature	0.00 °C
rel. humidity	0.00 %
Mass concentration	15.31 mg/m ³


Mask information

Mask ID	13868_20210104_142141
Mask los	M04-400010-29.12.2020-002
Mask type	TEST 1 / PALAS / FFP 2

Result

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 µm)	2.27 %	30.00 %	✓ pass
P_m	0.22 %		
P_Ph_45°	0.30 %	6.00 %	✓ pass

Distributions



Calibration valid until 05.11.2021