

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time
Jan Thomanek	05.01.2021	12:28:19

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.00 l/min
Generator air	2.97 l/min

Test conditions	
Temperature	0.00 °C
rel. humidity	0.00 %
Mass concentration	17.92 mg/m ³


Mask information

Mask ID	13868_20210105_122819
Mask los	050121-F02-001
Mask type	TEST 1 / PALAS / FFP 2

Result

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 µm)	1.65 %	30.00 %	✓ pass
P_m	0.21 %		
P_Ph_45°	0.35 %	6.00 %	✓ pass

Distributions



Calibration valid until 05.11.2021