

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time
Kübra Sabit	11.05.2021	09:56:21

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	25.30 °C
Aerosol	Pressurized air	rel. humidity	2.34 %

Mask information

Mask ID	13868_20210511_095621
Mask los	210504F02001-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.86 l/min	53.38 Pa	70.00 Pa	✓ pass
Inhalation 2	94.18 l/min	140.28 Pa	240.00 Pa	✓ pass
Exhalation	159.19 l/min	241.06 Pa	300.00 Pa	✓ pass



Calibration valid until 05.11.2021

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time
Kübra Sabit	11.05.2021	10:00:40

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.01 l/min
Generator air	2.97 l/min

Test conditions	
Temperature	25.35 °C
rel. humidity	4.19 %
Mass concentration	24.46 mg/m ³


Mask information

Mask ID	13868_20210511_100040
Mask los	210504F02001-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 µm)	1.22 %	30.00 %	✓ pass
P_m	0.07 %		
P_Ph_45°	0.16 %	6.00 %	✓ pass

Distributions



Calibration valid until 05.11.2021