

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time
Kübra Sabit	20.05.2021	10:32:18

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	22.81 °C
Aerosol	Pressurized air	rel. humidity	2.74 %

Mask information

Mask ID	13868_20210520_103218
Mask los	210519F01001-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.74 l/min	46.85 Pa	70.00 Pa	✔ pass
Inhalation 2	94.09 l/min	142.47 Pa	240.00 Pa	✔ pass
Exhalation	159.12 l/min	261.11 Pa	300.00 Pa	✔ pass



Calibration valid until 05.11.2021

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time
Kübra Sabit	20.05.2021	10:35:59

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.01 l/min
Generator air	2.97 l/min

Test conditions	
Temperature	22.81 °C
rel. humidity	4.20 %
Mass concentration	22.36 mg/m³


Mask information

Mask ID	13868_20210520_103559
Mask los	210519F01001-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 µm)	1.88 %	30.00 %	✓ pass
P_m	0.06 %		
P_Ph_45°	0.16 %	6.00 %	✓ pass

Distributions



Calibration valid until 05.11.2021