

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time
Demet Kilicaslan-Yalcin	05.07.2021	11:55:57

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	23.69 °C
Aerosol	Pressurized air	rel. humidity	2.73 %

Mask information

Mask ID	13868_20210705_115557
Mask los	210702F01001-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.88 l/min	44.51 Pa	70.00 Pa	✔ pass
Inhalation 2	94.24 l/min	135.48 Pa	240.00 Pa	✔ pass
Exhalation	159.09 l/min	236.65 Pa	300.00 Pa	✔ pass



Calibration valid until 05.11.2021

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time
Demet Kilicaslan-Yalcin	05.07.2021	12:00:35

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	93.50 l/min
Generator air	2.39 l/min

Test conditions	
Temperature	23.69 °C
rel. humidity	3.81 %
Mass concentration	90.89 mg/m ³


Mask information

Mask ID	13868_20210705_120035
Mask los	210702F01001-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 µm)	0.49 %	30.00 %	✓ pass
P_m	0.01 %		
P_Ph_45°	0.03 %	6.00 %	✓ pass

Distributions



Calibration valid until 05.11.2021