

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Svenja Wilk	14.09.2021	15:27:57	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	26.30 °C
Aerosol	Pressurized air	rel. humidity	3.04 %

Mask information

Mask ID	13868_20210914_152757
Mask lot	210907-F-L2-002-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.13 l/min	37.61 Pa	70.00 Pa	✔ pass
Inhalation 2	94.81 l/min	120.75 Pa	240.00 Pa	✔ pass
Exhalation	159.96 l/min	212.62 Pa	300.00 Pa	✔ pass

Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Svenja Wilk	14.09.2021	15:38:36	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.90 l/min
Generator air	2.53 l/min

Test conditions	
Temperature	26.42 °C
rel. humidity	2.14 %
Mass concentration	21.44 mg/m ³

Mask information


Mask ID	13868_20210914_153836
Mask lot	210907-F-L2-002-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	2.02 %	30.00 %	✓ pass
P_m	0.07 %		
P_Ph_45°	0.15 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022