

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Svenja Wilk	16.09.2021	16:20:52	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	25.05 °C
Aerosol	Pressurized air	rel. humidity	2.36 %

Mask information

Mask ID	13868_20210916_162052
Mask lot	210913-S-L2-002-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.07 l/min	37.12 Pa	70.00 Pa	✔ pass
Inhalation 2	94.86 l/min	118.68 Pa	240.00 Pa	✔ pass
Exhalation	160.09 l/min	211.49 Pa	300.00 Pa	✔ pass

Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Svenja Wilk	16.09.2021	16:28:16	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.89 l/min
Generator air	2.53 l/min

Test conditions	
Temperature	25.15 °C
rel. humidity	2.27 %
Mass concentration	18.95 mg/m ³

Mask information


Mask ID	13868_20210916_162816
Mask lot	210913-S-L2-002-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	1.90 %	30.00 %	✓ pass
P_m	0.06 %		
P_Ph_45°	0.14 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022