

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Christopher Reinhold	29.09.2021	15:40:08	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	22.67 °C
Aerosol	Pressurized air	rel. humidity	2.96 %

Mask information

Mask ID	13868_20210929_154008
Mask lot	210928-F-01-02-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.15 l/min	35.67 Pa	70.00 Pa	✔ pass
Inhalation 2	94.78 l/min	111.19 Pa	240.00 Pa	✔ pass
Exhalation	160.00 l/min	185.33 Pa	300.00 Pa	✔ pass

Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Christopher Reinhold	29.09.2021	15:44:44	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.96 l/min
Generator air	2.60 l/min

Test conditions	
Temperature	22.73 °C
rel. humidity	2.71 %
Mass concentration	9.58 mg/m ³

Mask information


Mask ID	13868_20210929_154444
Mask lot	210928-F-01-02-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	1.94 %	30.00 %	✓ pass
P_m	0.03 %		
P_Ph_45°	0.07 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022