

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Christopher Reinhold	01.10.2021	13:28:18	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	23.20 °C
Aerosol	Pressurized air	rel. humidity	3.42 %

Mask information

Mask ID	13868_20211001_132818
Mask lot	211001-F-01-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.02 l/min	30.79 Pa	70.00 Pa	✔ pass
Inhalation 2	94.80 l/min	98.23 Pa	240.00 Pa	✔ pass
Exhalation	159.95 l/min	172.70 Pa	300.00 Pa	✔ pass

Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Christopher Reinhold	01.10.2021	13:35:47	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	95.04 l/min
Generator air	2.84 l/min

Test conditions	
Temperature	23.13 °C
rel. humidity	2.83 %
Mass concentration	20.16 mg/m³

Mask information


Mask ID	13868_20211001_133547
Mask lot	211001-F-01-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	14.69 %	30.00 %	✓ pass
P_m	2.58 %		
P_Ph_45°	4.48 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022