

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Svenja Wilk	08.10.2021	15:31:41	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	24.86 °C
Aerosol	Pressurized air	rel. humidity	3.47 %

Mask information

Mask ID	13868_20211008_153141
Mask lot	211004-S-L3-002-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.18 l/min	30.60 Pa	70.00 Pa	✔ pass
Inhalation 2	94.80 l/min	96.33 Pa	240.00 Pa	✔ pass
Exhalation	160.03 l/min	166.57 Pa	300.00 Pa	✔ pass

Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Svenja Wilk	08.10.2021	15:45:27	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	95.04 l/min
Generator air	2.78 l/min

Test conditions	
Temperature	24.84 °C
rel. humidity	2.55 %
Mass concentration	18.86 mg/m³

Mask information


Mask ID	13868_20211008_154527
Mask lot	211004-S-L3-002-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	1.05 %	30.00 %	✓ pass
P_m	0.67 %		
P_Ph_45°	0.30 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022