

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Svenja Wilk	08.10.2021	16:56:40	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	25.74 °C
Aerosol	Pressurized air	rel. humidity	2.25 %

Mask information

Mask ID	13868_20211008_165640
Mask lot	211006-F-L1-002-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.17 l/min	31.49 Pa	70.00 Pa	✔ pass
Inhalation 2	94.89 l/min	99.74 Pa	240.00 Pa	✔ pass
Exhalation	160.01 l/min	170.92 Pa	300.00 Pa	✔ pass

Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Svenja Wilk	08.10.2021	17:00:52	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	95.01 l/min
Generator air	2.78 l/min

Test conditions	
Temperature	25.74 °C
rel. humidity	2.14 %
Mass concentration	22.47 mg/m³

Mask information


Mask ID	13868_20211008_170052
Mask lot	211006-F-L1-002-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	6.85 %	30.00 %	✓ pass
P_m	0.76 %		
P_Ph_45°	1.49 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022