

## Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Christopher Reinhold	29.10.2021	14:14:34	1.11.1.0

### Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	23.20 °C
Aerosol	Pressurized air	rel. humidity	3.57 %

### Mask information

Mask ID	13868_20211029_141434
Mask lot	211029-F-01-2-A-IN
Mask type	DMASKE / Chagentest / FFP 2

### Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.00 l/min	37.29 Pa	70.00 Pa	✔ pass
Inhalation 2	94.80 l/min	116.84 Pa	240.00 Pa	✔ pass
Exhalation	159.97 l/min	194.87 Pa	300.00 Pa	✔ pass

Calibration valid until 03.09.2022

# Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Christopher Reinhold	29.10.2021	14:19:36	1.11.1.0

## Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.89 l/min
Generator air	2.54 l/min

Test conditions	
Temperature	23.08 °C
rel. humidity	3.02 %
Mass concentration	16.98 mg/m³

## Mask information


Mask ID	13868_20211029_141936
Mask lot	211029-F-01-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

## Result

Calculation of the penetration is based on the following parameters:  $X_g = 0.29 \mu\text{m}$ ,  $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 $\mu\text{m}$ )	2.11 %	30.00 %	✓ pass
P_m	0.07 %		
P_Ph_45°	0.15 %	6.00 %	✓ pass

## Distributions



Calibration valid until 03.09.2022