

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Christopher Reinhold	09.11.2021	16:05:52	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	23.20 °C
Aerosol	Pressurized air	rel. humidity	3.03 %

Mask information

Mask ID	13868_20211109_160552
Mask lot	211109-F-01-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.09 l/min	33.52 Pa	70.00 Pa	✔ pass
Inhalation 2	94.77 l/min	99.75 Pa	240.00 Pa	✔ pass
Exhalation	159.93 l/min	164.62 Pa	300.00 Pa	✔ pass

Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Christopher Reinhold	09.11.2021	16:10:04	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.96 l/min
Generator air	2.59 l/min

Test conditions	
Temperature	23.20 °C
rel. humidity	2.64 %
Mass concentration	19.70 mg/m ³

Mask information


Mask ID	13868_20211109_161004
Mask lot	211109-F-01-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	3.32 %	30.00 %	✓ pass
P_m	0.10 %		
P_Ph_45°	0.19 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022