

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad Benchakroun	23.11.2021	08:48:53	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	21.25 °C
Aerosol	Pressurized air	rel. humidity	3.09 %

Mask information

Mask ID	13868_20211123_084853
Mask lot	211123-F-L4-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.02 l/min	32.87 Pa	70.00 Pa	✔ pass
Inhalation 2	94.78 l/min	96.25 Pa	240.00 Pa	✔ pass
Exhalation	159.97 l/min	164.88 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad Benchakroun	23.11.2021	08:54:12	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.97 l/min
Generator air	2.62 l/min

Test conditions	
Temperature	21.34 °C
rel. humidity	2.80 %
Mass concentration	17.92 mg/m³

Mask information


Mask ID	13868_20211123_085412
Mask lot	211123-F-L4-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	1.47 %	30.00 %	✓ pass
P_m	0.02 %		
P_Ph_45°	0.04 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022