

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad Benchakroun	01.12.2021	08:54:56	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	22.23 °C
Aerosol	Pressurized air	rel. humidity	2.87 %

Mask information

Mask ID	13868_20211201_085456
Mask lot	211127-F-L1-2-A
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.96 l/min	39.58 Pa	70.00 Pa	✔ pass
Inhalation 2	94.69 l/min	103.06 Pa	240.00 Pa	✔ pass
Exhalation	160.04 l/min	173.61 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad Benchakroun	01.12.2021	08:59:42	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.84 l/min
Generator air	2.69 l/min

Test conditions	
Temperature	22.30 °C
rel. humidity	2.64 %
Mass concentration	18.93 mg/m ³

Mask information


Mask ID	13868_20211201_085942
Mask lot	211127-F-L1-2-A
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	3.27 %	30.00 %	✓ pass
P_m	0.09 %		
P_Ph_45°	0.16 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022