

## Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Christopher Reinhold	02.12.2021	15:31:26	1.11.1.0

### Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	25.25 °C
Aerosol	Pressurized air	rel. humidity	2.82 %

### Mask information

Mask ID	13868_20211202_153126
Mask lot	211202-S-L4-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

### Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.10 l/min	31.94 Pa	70.00 Pa	✔ pass
Inhalation 2	94.83 l/min	97.34 Pa	240.00 Pa	✔ pass
Exhalation	159.99 l/min	179.87 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

## Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Christopher Reinhold	02.12.2021	15:36:43	1.11.1.0

### Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.89 l/min
Generator air	2.68 l/min

Test conditions	
Temperature	25.13 °C
rel. humidity	2.34 %
Mass concentration	22.85 mg/m <sup>3</sup>

### Mask information


Mask ID	13868_20211202_153643
Mask lot	211202-S-L4-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

### Result

Calculation of the penetration is based on the following parameters:  $X_g = 0.29 \mu\text{m}$ ,  $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 $\mu\text{m}$ )	2.44 %	30.00 %	✔ pass
P_m	0.21 %		
P_Ph_45°	0.27 %	6.00 %	✔ pass

### Distributions



Calibration valid until 03.09.2022