

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad Benchakroun	06.12.2021	10:07:30	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	18.88 °C
Aerosol	Pressurized air	rel. humidity	3.31 %

Mask information

Mask ID	13868_20211206_100730
Mask lot	211206-F-L3-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.04 l/min	36.93 Pa	70.00 Pa	✔ pass
Inhalation 2	94.81 l/min	104.40 Pa	240.00 Pa	✔ pass
Exhalation	159.88 l/min	176.34 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad Benchakroun	06.12.2021	10:15:48	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.89 l/min
Generator air	2.75 l/min

Test conditions	
Temperature	19.00 °C
rel. humidity	3.41 %
Mass concentration	17.32 mg/m ³

Mask information


Mask ID	13868_20211206_101548
Mask lot	211206-F-L3-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	2.19 %	30.00 %	✔ pass
P_m	0.19 %		
P_Ph_45°	0.17 %	6.00 %	✔ pass

Distributions



Calibration valid until 03.09.2022