

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad Benchakroun	07.12.2021	08:42:18	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	19.97 °C
Aerosol	Pressurized air	rel. humidity	2.93 %

Mask information

Mask ID	13868_20211207_084218
Mask lot	211207-F-L3-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.93 l/min	35.72 Pa	70.00 Pa	✔ pass
Inhalation 2	94.82 l/min	105.15 Pa	240.00 Pa	✔ pass
Exhalation	160.03 l/min	177.10 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad Benchakroun	07.12.2021	08:46:18	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.82 l/min
Generator air	2.94 l/min

Test conditions	
Temperature	20.03 °C
rel. humidity	2.90 %
Mass concentration	21.94 mg/m ³

Mask information


Mask ID	13868_20211207_084618
Mask lot	211207-F-L3-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	2.92 %	30.00 %	✔ pass
P_m	0.09 %		
P_Ph_45°	0.15 %	6.00 %	✔ pass

Distributions



Calibration valid until 03.09.2022