

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad Benchakroun	08.12.2021	08:31:43	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	22.12 °C
Aerosol	Pressurized air	rel. humidity	3.25 %

Mask information

Mask ID	13868_20211208_083143
Mask lot	211208-F-L1-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.00 l/min	34.44 Pa	70.00 Pa	✔ pass
Inhalation 2	94.61 l/min	103.44 Pa	240.00 Pa	✔ pass
Exhalation	159.97 l/min	177.45 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad Benchakroun	08.12.2021	08:38:32	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	95.02 l/min
Generator air	2.88 l/min

Test conditions	
Temperature	22.37 °C
rel. humidity	2.71 %
Mass concentration	21.64 mg/m³

Mask information


Mask ID	13868_20211208_083832
Mask lot	211208-F-L1-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	5.49 %	30.00 %	✔ pass
P_m	5.40 %		
P_Ph_45°	2.51 %	6.00 %	✔ pass

Distributions



Calibration valid until 03.09.2022