

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Asijat Gotschiyaeva	26.07.2022	12:20:36	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	27.72 °C
Aerosol	Pressurized air	rel. humidity	2.51 %

Mask information

Mask ID	13868_20220726_122036
Mask lot	220725-F-L5-2-MLK-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.08 l/min	23.47 Pa	70.00 Pa	✔ pass
Inhalation 2	94.85 l/min	86.42 Pa	240.00 Pa	✔ pass
Exhalation	159.98 l/min	165.63 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Asijat Gotschiyaeva	26.07.2022	12:26:26	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.85 l/min
Generator air	2.49 l/min

Test conditions	
Temperature	27.69 °C
rel. humidity	2.33 %
Mass concentration	23.11 mg/m³

Mask information


Mask ID	13868_20220726_122626
Mask lot	220725-F-L5-2-MLK-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	2.10 %	20.00 %	✓ pass
P_m	0.12 %		
P_Ph_45°	0.17 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022