

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Asijat Gotschiyaeva	28.07.2022	12:39:28	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	24.96 °C
Aerosol	Pressurized air	rel. humidity	3.45 %

Mask information

Mask ID	13868_20220728_123928
Mask lot	220728-F-L5-2-MLK-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.00 l/min	25.43 Pa	70.00 Pa	✔ pass
Inhalation 2	94.75 l/min	98.98 Pa	240.00 Pa	✔ pass
Exhalation	160.01 l/min	179.70 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Asijat Gotschiyaeva	28.07.2022	12:44:01	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.95 l/min
Generator air	2.59 l/min

Test conditions	
Temperature	25.15 °C
rel. humidity	2.71 %
Mass concentration	12.88 mg/m³

Mask information


Mask ID	13868_20220728_124401
Mask lot	220728-F-L5-2-MLK-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	2.73 %	20.00 %	✔ pass
P_m	0.18 %		
P_Ph_45°	0.21 %	6.00 %	✔ pass

Distributions



Calibration valid until 03.09.2022