

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Carsten	09.08.2022	12:33:40	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	26.23 °C
Aerosol	Pressurized air	rel. humidity	2.70 %

Mask information

Mask ID	13868_20220809_123340
Mask lot	220809-F-L5-2-LK-A-IN_DM
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.27 l/min	32.49 Pa	70.00 Pa	✔ pass
Inhalation 2	94.80 l/min	117.32 Pa	240.00 Pa	✔ pass
Exhalation	160.07 l/min	214.42 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Carsten	09.08.2022	12:37:52	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	95.02 l/min
Generator air	2.78 l/min

Test conditions	
Temperature	26.32 °C
rel. humidity	3.33 %
Mass concentration	19.97 mg/m³

Mask information


Mask ID	13868_20220809_123752
Mask lot	220809-F-L5-2-LK-A-IN_DM
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	5.72 %	20.00 %	✔ pass
P_m	8.76 %		
P_Ph_45°	2.02 %	6.00 %	✔ pass

Distributions



Calibration valid until 03.09.2022