

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Carsten	19.07.2022	13:30:18	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	29.35 °C
Aerosol	Pressurized air	rel. humidity	2.15 %

Mask information

Mask ID	13868_20220719_133018
Mask lot	220719-F-L5-1-LKbl-A-IN_D
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.33 l/min	40.53 Pa	70.00 Pa	✔ pass
Inhalation 2	94.82 l/min	152.09 Pa	240.00 Pa	✔ pass
Exhalation	160.05 l/min	293.65 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Carsten	19.07.2022	13:34:17	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.84 l/min
Generator air	2.43 l/min

Test conditions	
Temperature	29.45 °C
rel. humidity	2.72 %
Mass concentration	21.81 mg/m ³

Mask information


Mask ID	13868_20220719_133417
Mask lot	220719-F-L5-1-LKbl-A-IN_D
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	5.13 %	20.00 %	✔ pass
P_m	0.14 %		
P_Ph_45°	0.31 %	6.00 %	✔ pass

Distributions



Calibration valid until 03.09.2022