

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad Benchakroun	13.12.2021	15:07:01	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	23.92 °C
Aerosol	Pressurized air	rel. humidity	3.19 %

Mask information

Mask ID	13868_20211213_150701
Mask lot	211212-F-L3-2-M2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.01 l/min	31.22 Pa	70.00 Pa	✔ pass
Inhalation 2	94.77 l/min	95.49 Pa	240.00 Pa	✔ pass
Exhalation	159.90 l/min	167.34 Pa	300.00 Pa	✔ pass

Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad Benchakroun	13.12.2021	15:11:52	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.97 l/min
Generator air	2.63 l/min

Test conditions	
Temperature	23.93 °C
rel. humidity	2.71 %
Mass concentration	20.05 mg/m³

Mask information


Mask ID	13868_20211213_151152
Mask lot	211212-F-L3-2-M2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	2.67 %	30.00 %	✓ pass
P_m	0.40 %		
P_Ph_45°	0.55 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022