

## Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad Benchakroun	01.02.2022	11:56:40	1.11.1.0

### Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	22.71 °C
Aerosol	Pressurized air	rel. humidity	3.03 %

### Mask information

Mask ID	13868_20220201_115640
Mask lot	220201-F-L1-2-M2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

### Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.02 l/min	31.86 Pa	70.00 Pa	✔ pass
Inhalation 2	94.90 l/min	93.58 Pa	240.00 Pa	✔ pass
Exhalation	159.94 l/min	160.88 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

## Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad Benchakroun	01.02.2022	12:01:17	1.11.1.0

### Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.87 l/min
Generator air	2.68 l/min

Test conditions	
Temperature	22.73 °C
rel. humidity	2.56 %
Mass concentration	21.92 mg/m <sup>3</sup>

### Mask information


Mask ID	13868_20220201_120117
Mask lot	220201-F-L1-2-M2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

### Result

Calculation of the penetration is based on the following parameters:  $X_g = 0.29 \mu\text{m}$ ,  $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 $\mu\text{m}$ )	1.83 %	6.00 %	✓ pass
P_m	0.15 %		
P_Ph_45°	0.18 %	6.00 %	✓ pass

### Distributions



Calibration valid until 03.09.2022