

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad Benchakroun	04.04.2022	08:38:39	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	15.87 °C
Aerosol	Pressurized air	rel. humidity	4.00 %

Mask information

Mask ID	13868_20220404_083839
Mask lot	220330-S-L3-2-M2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.99 l/min	39.42 Pa	70.00 Pa	✔ pass
Inhalation 2	94.61 l/min	103.21 Pa	240.00 Pa	✔ pass
Exhalation	159.86 l/min	169.21 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad Benchakroun	04.04.2022	08:43:50	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.87 l/min
Generator air	2.75 l/min

Test conditions	
Temperature	15.80 °C
rel. humidity	3.56 %
Mass concentration	12.52 mg/m ³

Mask information


Mask ID	13868_20220404_084350
Mask lot	220330-S-L3-2-M2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	2.61 %	20.00 %	✓ pass
P_m	0.07 %		
P_Ph_45°	0.14 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022