

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Christopher Reinhold	09.12.2021	12:12:10	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	23.14 °C
Aerosol	Pressurized air	rel. humidity	3.11 %

Mask information

Mask ID	13868_20211209_121210
Mask lot	211209-F-L4-2-M2a-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.05 l/min	34.02 Pa	70.00 Pa	✔ pass
Inhalation 2	94.85 l/min	102.96 Pa	240.00 Pa	✔ pass
Exhalation	159.96 l/min	188.13 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Christopher Reinhold	09.12.2021	12:15:46	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	95.07 l/min
Generator air	3.13 l/min

Test conditions	
Temperature	23.20 °C
rel. humidity	4.29 %
Mass concentration	58.23 mg/m ³

Mask information


Mask ID	13868_20211209_121546
Mask lot	211209-F-L4-2-M2a-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	2.84 %	30.00 %	✓ pass
P_m	0.49 %		
P_Ph_45°	0.37 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022