

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Christopher Reinhold	27.12.2021	14:18:31	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	22.32 °C
Aerosol	Pressurized air	rel. humidity	3.77 %

Mask information

Mask ID	13868_20211227_141831
Mask lot	211227-F-L2-2-M2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.09 l/min	33.71 Pa	70.00 Pa	✔ pass
Inhalation 2	94.81 l/min	101.89 Pa	240.00 Pa	✔ pass
Exhalation	160.04 l/min	174.12 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Christopher Reinhold	27.12.2021	14:22:49	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.88 l/min
Generator air	2.68 l/min

Test conditions	
Temperature	22.18 °C
rel. humidity	3.12 %
Mass concentration	23.19 mg/m ³

Mask information


Mask ID	13868_20211227_142249
Mask lot	211227-F-L2-2-M2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	3.78 %	6.00 %	✓ pass
P_m	0.22 %		
P_Ph_45°	0.24 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022