

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad Benchakroun	11.04.2022	10:42:55	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	23.79 °C
Aerosol	Pressurized air	rel. humidity	2.22 %

Mask information

Mask ID	13868_20220411_104255
Mask lot	220411-F-L4-2-M2a-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.09 l/min	31.12 Pa	70.00 Pa	✔ pass
Inhalation 2	94.80 l/min	93.67 Pa	240.00 Pa	✔ pass
Exhalation	160.00 l/min	162.85 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad Benchakroun	11.04.2022	11:54:06	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	95.03 l/min
Generator air	2.84 l/min

Test conditions	
Temperature	24.77 °C
rel. humidity	2.05 %
Mass concentration	20.10 mg/m ³

Mask information


Mask ID	13868_20220411_115406
Mask lot	220411-F-L4-2-M2a-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	5.57 %	20.00 %	✓ pass
P_m	1.65 %		
P_Ph_45°	2.04 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022