

## Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Asijat Gotschiyaeva	07.06.2022	12:38:43	1.11.1.0

### Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	22.24 °C
Aerosol	Pressurized air	rel. humidity	9.96 %

### Mask information

Mask ID	13868_20220607_123843
Mask lot	220603-F-L4-2-M2a-A-IN
Mask type	DMASKE / Chargentest / FFP 2

### Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.20 l/min	29.83 Pa	70.00 Pa	✔ pass
Inhalation 2	94.69 l/min	101.05 Pa	240.00 Pa	✔ pass
Exhalation	160.02 l/min	179.38 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

# Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Asijat Gotschiyaeva	07.06.2022	12:48:44	1.11.1.0

## Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.98 l/min
Generator air	2.62 l/min

Test conditions	
Temperature	22.30 °C
rel. humidity	9.44 %
Mass concentration	19.09 mg/m³

## Mask information


Mask ID	13868_20220607_124844
Mask lot	220603-F-L4-2-M2a-A-IN
Mask type	DMASKE / Chargentest / FFP 2

## Result

Calculation of the penetration is based on the following parameters:  $X_g = 0.29 \mu\text{m}$ ,  $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 $\mu\text{m}$ )	2.71 %	20.00 %	✔ pass
P_m	0.71 %		
P_Ph_45°	0.29 %	6.00 %	✔ pass

## Distributions



Calibration valid until 03.09.2022