

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Asijat Gotschiyaeva	14.06.2022	12:55:55	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	22.81 °C
Aerosol	Pressurized air	rel. humidity	9.47 %

Mask information

Mask ID	13868_20220614_125555
Mask lot	220613-F-L3-2-M2a-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.13 l/min	25.00 Pa	70.00 Pa	✔ pass
Inhalation 2	94.85 l/min	86.46 Pa	240.00 Pa	✔ pass
Exhalation	159.99 l/min	153.48 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Asijat Gotschiyaeva	14.06.2022	13:00:17	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.88 l/min
Generator air	2.53 l/min

Test conditions	
Temperature	22.91 °C
rel. humidity	9.10 %
Mass concentration	18.48 mg/m ³

Mask information


Mask ID	13868_20220614_130017
Mask lot	220613-F-L3-2-M2a-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	3.28 %	20.00 %	✔ pass
P_m	0.08 %		
P_Ph_45°	0.18 %	6.00 %	✔ pass

Distributions



Calibration valid until 03.09.2022