

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Dustin	15.08.2022	16:21:33	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.88 l/min
Generator air	2.49 l/min

Test conditions	
Temperature	26.31 °C
rel. humidity	2.99 %
Mass concentration	18.60 mg/m ³

Mask information


Mask ID	13868_20220815_162133
Mask lot	220815-F-L4-2-M2a-A-IN_DM
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	3.01 %	20.00 %	✔ pass
P_m	0.09 %		
P_Ph_45°	0.20 %	6.00 %	✔ pass

Distributions



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Dustin	15.08.2022	16:17:18	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	26.23 °C
Aerosol	Pressurized air	rel. humidity	2.21 %

Mask information

Mask ID	13868_20220815_161718
Mask lot	220815-F-L4-2-M2a-A-IN_DM
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.14 l/min	22.94 Pa	70.00 Pa	✔ pass
Inhalation 2	94.90 l/min	89.53 Pa	240.00 Pa	✔ pass
Exhalation	160.03 l/min	159.75 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022