

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Dustin	16.08.2022	13:21:22	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.84 l/min
Generator air	2.49 l/min

Test conditions	
Temperature	25.43 °C
rel. humidity	3.42 %
Mass concentration	20.44 mg/m³

Mask information


Mask ID	13868_20220816_132122
Mask lot	220815-F-L4-2-M2a-A-IN_DM
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	3.26 %	20.00 %	✔ pass
P_m	0.75 %		
P_Ph_45°	0.26 %	6.00 %	✔ pass

Distributions



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Dustin	16.08.2022	13:17:10	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	25.37 °C
Aerosol	Pressurized air	rel. humidity	2.63 %

Mask information

Mask ID	13868_20220816_131710
Mask lot	220815-F-L4-2-M2a-A-IN_DM
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.23 l/min	28.57 Pa	70.00 Pa	✔ pass
Inhalation 2	94.77 l/min	101.58 Pa	240.00 Pa	✔ pass
Exhalation	159.95 l/min	180.09 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022