

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Hikmet	15.11.2022	14:30:30	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.80 l/min
Generator air	2.68 l/min

Test conditions	
Temperature	21.31 °C
rel. humidity	3.69 %
Mass concentration	23.39 mg/m³

Mask information


Mask ID	13868_20221115_143030
Mask lot	221114-F-L3-M2A-A-IN_
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	3.12 %	20.00 %	✔ pass
P_m	0.06 %		
P_Ph_45°	0.15 %	6.00 %	✔ pass

Distributions



Calibration expired since 03.09.2022

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Hikmet	15.11.2022	14:23:43	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	21.24 °C
Aerosol	Pressurized air	rel. humidity	4.53 %

Mask information

Mask ID	13868_20221115_142343
Mask lot	221114-F-L3-M2A-A-IN_
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.99 l/min	29.48 Pa	70.00 Pa	✔ pass
Inhalation 2	94.80 l/min	101.51 Pa	240.00 Pa	✔ pass
Exhalation	159.98 l/min	174.11 Pa	300.00 Pa	✔ pass



Calibration expired since 03.09.2022