

## Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Hikmet	22.11.2022	10:41:17	1.11.1.0

### Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.92 l/min
Generator air	2.74 l/min

Test conditions	
Temperature	22.81 °C
rel. humidity	2.71 %
Mass concentration	25.96 mg/m <sup>3</sup>

### Mask information


Mask ID	13868_20221122_104117
Mask lot	221121-F-L3-M2A-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

### Result

Calculation of the penetration is based on the following parameters:  $X_g = 0.29 \mu\text{m}$ ,  $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 $\mu\text{m}$ )	3.38 %	20.00 %	✓ pass
P_m	0.10 %		
P_Ph_45°	0.21 %	6.00 %	✓ pass

### Distributions



Calibration expired since 03.09.2022

## Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Hikmet	22.11.2022	10:37:37	1.11.1.0

### Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	22.81 °C
Aerosol	Pressurized air	rel. humidity	2.95 %

### Mask information

Mask ID	13868_20221122_103737
Mask lot	221121-F-L3-M2A-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

### Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.14 l/min	28.78 Pa	70.00 Pa	✔ pass
Inhalation 2	94.77 l/min	91.44 Pa	240.00 Pa	✔ pass
Exhalation	159.99 l/min	159.15 Pa	300.00 Pa	✔ pass



Calibration expired since 03.09.2022