

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Hikmet	23.11.2022	11:13:30	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.98 l/min
Generator air	2.62 l/min

Test conditions	
Temperature	21.17 °C
rel. humidity	3.02 %
Mass concentration	17.32 mg/m³

Mask information


Mask ID	13868_20221123_111330
Mask lot	221123-F-L4-M2A-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	2.68 %	20.00 %	✓ pass
P_m	0.06 %		
P_Ph_45°	0.13 %	6.00 %	✓ pass

Distributions



Calibration expired since 03.09.2022

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Hikmet	23.11.2022	11:09:53	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	21.12 °C
Aerosol	Pressurized air	rel. humidity	3.03 %

Mask information

Mask ID	13868_20221123_110953
Mask lot	221123-F-L4-M2A-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.09 l/min	29.79 Pa	70.00 Pa	✔ pass
Inhalation 2	94.79 l/min	95.01 Pa	240.00 Pa	✔ pass
Exhalation	159.94 l/min	164.27 Pa	300.00 Pa	✔ pass



Calibration expired since 03.09.2022