

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Dustin	18.11.2022	11:49:02	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.86 l/min
Generator air	2.68 l/min

Test conditions	
Temperature	22.12 °C
rel. humidity	2.74 %
Mass concentration	19.91 mg/m³

Mask information


Mask ID	13868_20221118_114902
Mask lot	221116-F-L4-M2B-2-A-IN_DM
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	4.79 %	20.00 %	✔ pass
P_m	0.07 %		
P_Ph_45°	0.16 %	6.00 %	✔ pass

Distributions



Calibration expired since 03.09.2022

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Dustin	18.11.2022	11:44:52	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	22.12 °C
Aerosol	Pressurized air	rel. humidity	3.03 %

Mask information

Mask ID	13868_20221118_114452
Mask lot	221116-F-L4-M2B-2-A-IN_DM
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.03 l/min	30.92 Pa	70.00 Pa	✔ pass
Inhalation 2	94.77 l/min	102.37 Pa	240.00 Pa	✔ pass
Exhalation	160.02 l/min	178.66 Pa	300.00 Pa	✔ pass



Calibration expired since 03.09.2022