

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Hikmet	22.11.2022	10:53:54	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.91 l/min
Generator air	2.68 l/min

Test conditions	
Temperature	23.08 °C
rel. humidity	2.56 %
Mass concentration	23.65 mg/m ³

Mask information


Mask ID	13868_20221122_105354
Mask lot	221121-F-L4-M2B-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	1.92 %	20.00 %	✔ pass
P_m	0.06 %		
P_Ph_45°	0.15 %	6.00 %	✔ pass

Distributions



Calibration expired since 03.09.2022

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Hikmet	22.11.2022	10:50:20	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	23.00 °C
Aerosol	Pressurized air	rel. humidity	2.94 %

Mask information

Mask ID	13868_20221122_105020
Mask lot	221121-F-L4-M2B-2-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.17 l/min	33.57 Pa	70.00 Pa	✔ pass
Inhalation 2	94.78 l/min	107.65 Pa	240.00 Pa	✔ pass
Exhalation	160.00 l/min	188.56 Pa	300.00 Pa	✔ pass



Calibration expired since 03.09.2022