

## Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad Benchakroun	03.05.2022	10:13:03	1.11.1.0

### Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	22.21 °C
Aerosol	Pressurized air	rel. humidity	3.54 %

### Mask information

Mask ID	13868_20220503_101303
Mask lot	220429-S-L4-2-2b-RT-A-IN
Mask type	DMASKE / Chargentest / FFP 2

### Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.03 l/min	32.99 Pa	70.00 Pa	✔ pass
Inhalation 2	94.84 l/min	100.34 Pa	240.00 Pa	✔ pass
Exhalation	159.92 l/min	172.15 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

# Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad Benchakroun	03.05.2022	10:37:17	1.11.1.0

## Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.99 l/min
Generator air	2.63 l/min

Test conditions	
Temperature	22.20 °C
rel. humidity	3.13 %
Mass concentration	16.52 mg/m³

## Mask information


Mask ID	13868_20220503_103717
Mask lot	220429-S-L4-2-2b-RT-A-IN
Mask type	DMASKE / Chargentest / FFP 2

## Result

Calculation of the penetration is based on the following parameters:  $X_g = 0.29 \mu\text{m}$ ,  $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 $\mu\text{m}$ )	1.29 %	20.00 %	✓ pass
P_m	0.05 %		
P_Ph_45°	0.10 %	6.00 %	✓ pass

## Distributions



Calibration valid until 03.09.2022