

## Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad	03.05.2022	09:19:29	1.11.1.0

### Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	21.34 °C
Aerosol	Pressurized air	rel. humidity	3.21 %

### Mask information

Mask ID	13868_20220503_091929
Mask lot	220428-F-L4-2-2b-GR-A-IN
Mask type	DMASKE / Chargentest / FFP 2

### Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.99 l/min	40.72 Pa	70.00 Pa	✔ pass
Inhalation 2	94.72 l/min	113.08 Pa	240.00 Pa	✔ pass
Exhalation	159.95 l/min	187.03 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

# Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad	03.05.2022	09:23:19	1.11.1.0

## Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.94 l/min
Generator air	2.59 l/min

Test conditions	
Temperature	21.34 °C
rel. humidity	3.03 %
Mass concentration	17.33 mg/m³

## Mask information


Mask ID	13868_20220503_092319
Mask lot	220428-F-L4-2-2b-GR-A-IN
Mask type	DMASKE / Chargentest / FFP 2

## Result

Calculation of the penetration is based on the following parameters:  $X_g = 0.29 \mu\text{m}$ ,  $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 $\mu\text{m}$ )	1.94 %	20.00 %	✓ pass
P_m	0.29 %		
P_Ph_45°	0.46 %	6.00 %	✓ pass

## Distributions



Calibration valid until 03.09.2022