

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Demet	23.03.2022	14:52:38	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	24.18 °C
Aerosol	Pressurized air	rel. humidity	2.34 %

Mask information

Mask ID	13868_20220323_145238
Mask lot	220323-F-L5-2-LK-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.08 l/min	35.12 Pa	70.00 Pa	✔ pass
Inhalation 2	94.80 l/min	102.67 Pa	240.00 Pa	✔ pass
Exhalation	159.96 l/min	169.11 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Demet	23.03.2022	14:56:59	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.98 l/min
Generator air	2.62 l/min

Test conditions	
Temperature	24.20 °C
rel. humidity	2.14 %
Mass concentration	19.53 mg/m ³

Mask information


Mask ID	13868_20220323_145659
Mask lot	220323-F-L5-2-LK-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	1.63 %	20.00 %	✓ pass
P_m	0.07 %		
P_Ph_45°	0.14 %	6.00 %	✓ pass

Distributions



Calibration valid until 03.09.2022