

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Dustin	26.09.2022	10:38:21	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.84 l/min
Generator air	2.68 l/min

Test conditions	
Temperature	23.96 °C
rel. humidity	3.80 %
Mass concentration	23.33 mg/m³

Mask information


Mask ID	13868_20220926_103821
Mask lot	220923-F-L5-2-MLK-A-IN_DM
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	5.15 %	20.00 %	✔ pass
P_m	0.19 %		
P_Ph_45°	0.31 %	6.00 %	✔ pass

Distributions



Calibration expired since 03.09.2022

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Dustin	26.09.2022	10:34:37	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	23.95 °C
Aerosol	Pressurized air	rel. humidity	3.44 %

Mask information

Mask ID	13868_20220926_103437
Mask lot	220923-F-L5-2-MLK-A-IN_DM
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.21 l/min	33.39 Pa	70.00 Pa	✔ pass
Inhalation 2	94.73 l/min	123.82 Pa	240.00 Pa	✔ pass
Exhalation	159.97 l/min	252.14 Pa	300.00 Pa	✔ pass



Calibration expired since 03.09.2022