

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Dustin	24.10.2022	12:48:30	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.90 l/min
Generator air	2.53 l/min

Test conditions	
Temperature	24.59 °C
rel. humidity	2.32 %
Mass concentration	20.14 mg/m ³

Mask information


Mask ID	13868_20221024_124830
Mask lot	221020-F-L5-2-MLK-A-IN_D_
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	2.21 %	20.00 %	✔ pass
P_m	0.14 %		
P_Ph_45°	0.21 %	6.00 %	✔ pass

Distributions



Calibration expired since 03.09.2022

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Dustin	24.10.2022	12:44:59	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	24.62 °C
Aerosol	Pressurized air	rel. humidity	2.98 %

Mask information

Mask ID	13868_20221024_124459
Mask lot	221020-F-L5-2-MLK-A-IN_D_
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.02 l/min	37.83 Pa	70.00 Pa	✔ pass
Inhalation 2	94.81 l/min	131.33 Pa	240.00 Pa	✔ pass
Exhalation	160.03 l/min	245.30 Pa	300.00 Pa	✔ pass



Calibration expired since 03.09.2022