

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Dustin	31.10.2022	09:38:55	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.87 l/min
Generator air	2.68 l/min

Test conditions	
Temperature	23.25 °C
rel. humidity	2.74 %
Mass concentration	19.45 mg/m³

Mask information


Mask ID	13868_20221031_093855
Mask lot	221026-F-L5-2-MLK-A-IN_D_
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	1.95 %	20.00 %	✔ pass
P_m	0.06 %		
P_Ph_45°	0.14 %	6.00 %	✔ pass

Distributions



Calibration expired since 03.09.2022

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Dustin	31.10.2022	09:35:17	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	23.27 °C
Aerosol	Pressurized air	rel. humidity	3.24 %

Mask information

Mask ID	13868_20221031_093517
Mask lot	221026-F-L5-2-MLK-A-IN_D_
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.06 l/min	47.34 Pa	70.00 Pa	✔ pass
Inhalation 2	94.78 l/min	137.76 Pa	240.00 Pa	✔ pass
Exhalation	160.04 l/min	256.58 Pa	300.00 Pa	✔ pass



Calibration expired since 03.09.2022