

## Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad Benchakroun	07.12.2021	10:39:35	1.11.1.0

### Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	21.34 °C
Aerosol	Pressurized air	rel. humidity	3.14 %

### Mask information

Mask ID	13868_20211207_103935
Mask lot	211205-F-L1-2-XS-A-IN
Mask type	DMASKE / Chargentest / FFP 2

### Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.99 l/min	33.45 Pa	70.00 Pa	✔ pass
Inhalation 2	94.70 l/min	103.45 Pa	240.00 Pa	✔ pass
Exhalation	159.97 l/min	174.80 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

## Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad Benchakroun	07.12.2021	10:57:12	1.11.1.0

### Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.83 l/min
Generator air	2.93 l/min

Test conditions	
Temperature	21.54 °C
rel. humidity	2.64 %
Mass concentration	22.40 mg/m <sup>3</sup>

### Mask information


Mask ID	13868_20211207_105712
Mask lot	211205-F-L2-2-XS-A-IN
Mask type	DMASKE / Chargentest / FFP 2

### Result

Calculation of the penetration is based on the following parameters:  $X_g = 0.29 \mu\text{m}$ ,  $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 $\mu\text{m}$ )	5.38 %	30.00 %	✓ pass
P_m	4.08 %		
P_Ph_45°	4.31 %	6.00 %	✓ pass

### Distributions



Calibration valid until 03.09.2022