

Test report - PMFT 1000 - Breath resistance / Valve test

Operator	Date	Time	Vesion
Saad Benchakroun	15.12.2021	09:28:39	1.11.1.0

Basic information

Operation parameter		Test conditions	
Script	Atemwiderstand	Temperature	22.03 °C
Aerosol	Pressurized air	rel. humidity	3.82 %

Mask information

Mask ID	13868_20211215_092839
Mask lot	211215-F-L2-1-MXS-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	30.01 l/min	29.27 Pa	70.00 Pa	✔ pass
Inhalation 2	94.77 l/min	91.14 Pa	240.00 Pa	✔ pass
Exhalation	159.92 l/min	192.81 Pa	300.00 Pa	✔ pass



Calibration valid until 03.09.2022

Test report - PMFT 1000 - Oil penetration

Operator	Date	Time	Vesion
Saad Benchakroun	15.12.2021	09:34:35	1.11.1.0

Basic information

Operation parameter	
Script	Effizienz Test (Öl)
Aerosol	Oil
Main air	94.98 l/min
Generator air	2.63 l/min

Test conditions	
Temperature	22.02 °C
rel. humidity	3.03 %
Mass concentration	19.24 mg/m ³

Mask information


Mask ID	13868_20211215_093435
Mask lot	211215-F-L2-1-MXS-A-IN
Mask type	DMASKE / Chargentest / FFP 2

Result

Calculation of the penetration is based on the following parameters: $X_g = 0.29 \mu\text{m}$, $\text{Sigma} = 1.85$

Penetration	Measurement	Upper limit	Result
P_SARS_CoV-2 (0.15 μm)	3.05 %	30.00 %	✔ pass
P_m	1.98 %		
P_Ph_45°	2.20 %	6.00 %	✔ pass

Distributions



Calibration valid until 03.09.2022